

Bone mineralizing hormone <calcitonin> produced in cartilaginous fish : clarification of its physiological function in cellular level

メタデータ	言語: jpn 出版者: 公開日: 2022-06-09 キーワード (Ja): キーワード (En): 作成者: Sasayama, Yuichi メールアドレス: 所属:
URL	https://doi.org/10.24517/00066214

This work is licensed under a Creative Commons
Attribution-NonCommercial-ShareAlike 3.0
International License.



1996 Fiscal Year Final Research Report Summary

Bone mineralizing hormone <calcitonin> produced in cartilaginous fish : clarification of its physiological function in cellular level

Research Project

Project/Area Number

07554089

Research Category

Grant-in-Aid for Scientific Research (A)

Allocation Type

Single-year Grants

Section

試験

Research Field

生物形態・構造

Research Institution

Kanazawa University

Principal Investigator

SASAYAMA Yuichi Kanazawa University Faculty of science Professor, 理学部, 教授 (30018999)

Co-Investigator(Kenkyū-buntansha)

SUZUKI Nobuo Kanazawa University Faculty of Science Assistant Prof., 理学部, 助手 (60242476)

NISHIZAWA Tsutomu Instit.of Sumitomo Hi Qual.Life Chief Investigator, 生物工学研, 室長

KURODA Hideyo Toyama University Faculty of Science Professor, 理学部, 教授 (50064845)

Project Period (FY)

1995 - 1996

Keywords

Research Abstract

As we have clarified a primary structure of calcitonin from the stingray before, as a next step, the physiological function of this hormone was tried to study at a cellular level. This project was performed in 1995 and 1996. At present, however, the fruitful results are not yet obtained for some reasons. The main difficulty was to have to build up the instrument for determining cytoplasmic Ca level by ourselves, because financial support was fairly small when compared to the amount requested. Nevertheless, we made an instrument for the purpose by combining an inversed microscope (Nikon TMD 300) with an apparatus of epi-fluorescence for Indo-1 (Nikon EF-S/Ca) and 2 photo multipliers (Hamamatsu Photonics R1477). The principle of this instrument is following. When the cell is irradiated by a wave length 380 nm from a mercury lamp, the fluorescence must be released from the cell, because the cell is treated with Indo-1 already.

The fluorescence is divided into two wave lengths of 405 nm and 485 nm by a dichroic mirror DM455. If Ca level in the cytoplasm is high, the peak of 405 nm should be high, and the Ca level is low, peak of 485 nm should be high. Such changes in the Ca level is recorded by a two-pen recorder. On the other hand, although we tried to cultivate chondrocyte from branchial cartilage of goldfish in vitro, but good results are not yet obtained. Recently, it was reported that cellular membrane of neurons is composed of the same component as the cartilaginous matrix of cartilage. Therefore, to check the performance of our instrument, we examined the cytoplasmic changes of Ca level in neurons from abdominal ganglion of Aplysia, as we have checked some of neurons can respond to calcitonin before. However, we could not detect Ca changes, when calcitonin was added to the incubation medium. At present, the causes are examined, whether the sensitivity of this instrument was low, or quantity of the administration of calcitonin was too much low.

URL: https://kaken.nii.ac.jp/report/KAKENHI-PROJECT-07554089/075540891996kenkyu_seika_hokoku_

Published: 1999-03-08