

Temporal Attenuation of Iodine Content and its Effect on the Antibacterial Activity of Iodine-Supported Titanium Implants

Takashi Kato¹, Toshiharu Shirai¹, Norio Yamamoto^{1*}, Hideji Nishida¹, Katsuhiko Hayashi¹, Akihiko Takeuchi¹, Shinji Miwa¹, Kaori Ohtani² and Hiroyuki Tsuchiya¹

¹Department of Orthopaedic Surgery, Graduate School of Medical Science, Kanazawa University, 13-1 Takaramachi Kanazawa, Japan

²Department of Bacteriology, Kanazawa University, 13-1 Takaramachi Kanazawa, Japan

Abstract

Infections related to orthopedic implants often require prolonged therapy and complex interventions; accordingly, the development of implants with a low risk of infection is a high priority. Iodine-supported titanium implants with antibacterial activity are safe and effective for prophylaxis against implant-related infections. However, temporal changes in the iodine content of implants and the relationship between iodine content and antibacterial activity have not been investigated. Temporal changes in iodine for 10–12 $\mu\text{g}/\text{cm}^2$ iodine-supplemented titanium implants were investigated *in vitro* and *in vivo* using rabbit models (subcutaneous soft tissue, intra-articular, and endo-osseous sites). *In vitro* antibacterial activity against *Staphylococcus aureus* and *E. coli* were also investigated using implants with various iodine contents (0%, 20%, 50%, 60%, and 100%, where 100% corresponds to 13 $\mu\text{g}/\text{cm}^2$ iodine based on current implants in clinical use). The residual iodine after 1 year of implantation and the minimum effective iodine concentration required for antibacterial activity were determined. *In vitro* and *in vivo* experiments showed a similar temporal pattern of initially rapid and subsequently slow attenuation of iodine in the implants, with approximately 30% of the initial iodine content remaining at 1 year. Pure titanium implants and implants with a 0% oxide layer did not exhibit antibacterial activity. Titanium implants supplemented with 20%, 50%, 80%, and 100% iodine showed *in vitro* antibacterial activity that varied in a dose-dependent and duration-dependent manner. Implants with $\geq 20\%$ iodine achieved complete clearance of *S. aureus* and *E. coli* colonies by 24 hours of incubation. Implants with iodine contents of $\geq 20\%$ demonstrated sufficient antibacterial activity, indicating that current iodine-supported titanium implants possess adequate antibacterial activity to prevent implant-related infections, even after 1 year of implantation. These results support the clinical use of iodine-supported titanium implants to prevent orthopedic implant-related infections.

Keywords: Iodine-supported titanium, Antibacterial activity, Temporal attenuation, Implant infection

Abbreviations: *E. coli*: *Escherichia coli*

PBS: phosphate-buffered saline

S. aureus: *Staphylococcus aureus*

Introduction

A wide variety of implants is currently used in orthopedic procedures. Most conventional implants do not possess antimicrobial activity and implant-related infections continue to be a significant complication. Implant-related infections occur in 0.7% to 4.2% of patients after orthopedic surgery and are a particular concern after prosthetic replacement arthroplasties [1-5]. The risk of infection is particularly high in patients receiving chemotherapy for orthopedic conditions, such as bone soft tissue tumors and rheumatoid arthritis. Indeed, the infection rate is 1–2% in patients with total knee arthroplasty, but 14.6–17% for cases of bone soft tissue tumors of the knee that underwent reconstruction with implants [6-10].

The treatment of orthopedic implant-related infections is challenging because bone infections require the long-term administration of antibacterial agents, irrigation, multiple surgeries, and implant removal or replacement. The prevention of such infections is therefore very important. Various biomaterial surface modifications of stainless steel and titanium orthopedic implants have been developed for prophylaxis against implant-related infections. The covalent attachment of polycationic groups, implantation of Ca^+ , N^+ , and F^+ ions with antimicrobial agents, and alloying and surface processing of silver have been investigated. We have developed a technique to form a highly porous anodic oxide film on the surface of titanium implants and impregnate iodine into the pores; this strategy imparts antimicrobial

activity to the implants [11-19]. In general, iodine-supported titanium implants have a 5–10- μm -thick oxide film with $>50,000$ pores/ mm^2 loaded with 10–12 $\mu\text{g}/\text{cm}^2$ iodine. In our previous studies, we have demonstrated the *in vitro* antibacterial activity, *in vitro* cytocompatibility, and *in vivo* safety characteristics of iodine-supported titanium implants [18]. Early results from ongoing clinical trials have also demonstrated the safety and effectiveness of these implants in patients [20-23]. We have observed that the iodine content on the implant surface decreases over time; however, temporal changes in the iodine content and the relationship between iodine content and antibacterial activity have not been investigated. The purpose of this study was to examine the time course of the iodine content using *in vitro* and *in vivo* experimental models and to demonstrate the relationship between iodine content and antibacterial activity.

Materials and Methods

Study design

Iodine-supported titanium implants were produced in various shapes and with different quantities of iodine. *in vitro* antibacterial

*Corresponding author: Yamamoto N, Department of Orthopaedic Surgery, Graduate School of Medical Science, Kanazawa University, 13-1 Takaramachi Kanazawa, Japan, Tel: +81-76-265-2374; E-mail: norinori@med.kanazawa-u.ac.jp

Received May 25, 2016; Accepted June 28, 2016; Published July 08, 2016

Citation: Kato T, Shirai T, Yamamoto N, Nishida H, Hayashi K, et al. (2016) Temporal Attenuation of Iodine Content and its Effect on the Antibacterial Activity of Iodine-Supported Titanium Implants. J Microb Biochem Technol 8: 285-289. doi: [10.4172/1948-5948.1000298](http://dx.doi.org/10.4172/1948-5948.1000298)

Copyright: © 2016 Kato T, et al. This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

activity against *S. aureus* and *E. coli* was investigated using implants with different iodine contents to determine the minimum effective iodine content required for antibacterial activity. The temporal change in iodine content in the implants over 1 year was also investigated using *in vitro* and *in vivo* rabbit models.

Iodine-supported implants

Iodine-supported titanium implants were produced by Prostec Instruments Company (Sabae, Japan) using a technique described previously [18,19]. Briefly, a 5–10- μm -thick anodic oxide film with >50,000 pores/ mm^2 was formed on the surface of titanium implants. Ionized iodine was electrodeposited within these pores to achieve the desired iodine content (0–13 $\mu\text{g}/\text{cm}^2$ iodine). Four types of iodine-supported titanium implants were developed as follows: (1) square plates (20 mm \times 20 mm \times 2 mm); (2) circular discs (6 mm diameter, 2 mm thick); (3) washers (outer diameter 16 mm, inner diameter 6.5 mm, 2 mm thick); and (4) screws (6.5 mm diameter, 16 mm length). The iodine content on implants was measured by X-ray fluorescence spectroscopy (Prostec Instruments Company).

In vitro temporal assessment of iodine contents in implants

Square implants with iodine contents of 10–12 $\mu\text{g}/\text{cm}^2$ ($n = 3$ per condition) were immersed in 30 mL of PBS in a 50-mL centrifuge tube and incubated at 37°C for various durations up to 1 year. After the specified periods (3 hours, 12 hours, 24 hours, 10 days, 50 days, and 1 year), the implants were removed and the iodine content was measured by X-ray fluorescence spectroscopy (Prostec Instruments Company) and averaged for the three implants. Residual iodine is expressed as a percentage of the iodine content in the fresh implant.

In vivo temporal assessment of iodine contents in implants

A previously described rabbit model was used for the *in vivo* analysis [18]. All animal experiments were conducted with the approval of the Institutional Animal Care and Use Committee and carried out in strict accordance with its regulations. Mature female Japanese white rabbits ($n = 15$) weighing 2.5–3.0 kg were anesthetized by an intramuscular injection of ketamine hydrochloride (50 mg/kg; Warner-Lambert,

Morris Plains, NJ, USA) and an intravenous injection of pentobarbital sodium (40–50 mg/kg). Each animal received three iodine-supported titanium implants (one circular disc, one washer, and one screw; iodine content, 10–12 $\mu\text{g}/\text{cm}^2$) at three different anatomical sites (subcutaneous soft tissue, intra-articular, and endo-osseous) relevant for orthopedic implants (Figure 1).

For the subcutaneous soft tissue model, a circular disc was implanted in the dorsal aspect of the rabbit. The knee joint was used to model the joint and bone sites. A screw was inserted through a washer into the distal intra-articular end of the femur, with the screw and washer representing the bone and joint site, respectively (Figure 1). At specific postoperative periods (14, 28, 90, 200, and 365 days), the animals were killed ($n=3$ per time point), implants were retrieved, and residual iodine was measured by X-ray fluorescence spectroscopy (Prostec Instruments Company). Values obtained for three animals at each time point were averaged for the analysis.

In vitro antibacterial activity

Circular disc implants (diameter: 20 mm; thickness: 2 mm) were used. Five types of implants were fabricated with different iodine content: 0 $\mu\text{g}/\text{cm}^2$ (0% oxide layer), 3 $\mu\text{g}/\text{cm}^2$ (20%), 7 $\mu\text{g}/\text{cm}^2$ (50%), 10 $\mu\text{g}/\text{cm}^2$ (80%), and 13 $\mu\text{g}/\text{cm}^2$ (100%). Note that 13 $\mu\text{g}/\text{cm}^2$ was considered 100% because it is the standard amount in clinical implants [18]. Pure titanium implants were used as controls.

The implants were exposed to gram-positive *S. aureus* strain 25923 (ATCC, Manassas, VA, USA) or gram-negative *E. coli* strain MG1655 (ATCC). The antibacterial activity of the implants was measured using a method approved by Japanese Industrial Standards as previously described [18]. Approximately 10^6 colony-forming units were inoculated on an autoclaved circular implant placed in a sterile glass Petri dish, which was then covered and incubated at 37°C for 2, 6, 24, or 48 hours ($n = 3$ per implant type). At each time point, each implant was washed with 5 mL of phosphate-buffered saline (PBS). The washed eluate was diluted 1:50 with PBS and 100 μL of the dilute eluate was incubated in Brain Heart Infusion broth for *S. aureus* and LB broth (1% w/v tryptone, 0.5% w/v yeast extract, 0.5% w/v NaCl) for *E. coli* at 37°C. The number of bacterial colonies was counted after 24 hours and the average number of colonies for the triplicate implants at different incubation periods (2–48 hours) was used to determine antibacterial activity against *S. aureus* and *E. coli* (Figure 2).




Figure 1: Study design for *in vivo* temporal assessments of the iodine content in implants. Each rabbit received three implants—a subcutaneous circular implant, an intra-articular washer, and an endo-osseous screw. Animals were euthanized at various time points over 1 year and implants were retrieved to measure the residual iodine




Figure 2: Study design for *in vitro* antibacterial activity assessment using a modified version of the Japanese Industrial Standards method.

For the statistical analysis, Student's t-tests were performed for parametric tests and differences among means were considered statistically significant when $p < 0.05$.

Results

In vitro temporal assessment of iodine quantities in implants

Residual iodine in the implants decreased gradually over the duration of the *in vitro* study. The average residual iodine was 68.7% at 3 hours, 62.7% at 12 hours, 53.5% at 24 hours, 41.6% at 10 days, 24.8% at 50 days, and 29.8% at 365 days (Figure 3).

In vivo temporal assessment of iodine quantities in implants

Residual iodine in the implants also decreased gradually over time in the *in vivo* study at all three anatomical locations (Figure 4).

In the subcutaneous soft tissue (circular implant), the average residual iodine was 59.7% at 14 days, 54.4% at 28 days, 34.1% at 90 days, 23.1% at 200 days, and 27.3% at 365 days. In the intra-articular location (washer implant), the average residual iodine was 54.7% at 14 days, 51.1% at 28 days, 53.7% at 90 days, 42.7% at 200 days, and 31.5% at 365 days. Similarly, in the intra-osseous location (screw implant), the




Figure 3: *In vitro* temporal profile of iodine content in implants, showing an initially rapid attenuation, followed by a more gradual attenuation during incubation in phosphate-buffered saline (PBS). Implants retained approximately 30% of the initial iodine content at the end of 1 year.

The vertical axis is for iodine content.




Figure 4: *In vivo* temporal profile of iodine content in implants, showing an initially rapid attenuation, followed by a more gradual attenuation at all three anatomical sites. All implants retained approximately 30% of the initial iodine content at the end of 1 year.

The vertical axis is for iodine content.

average residual iodine was 47.4% at 14 days, 47.1% at 28 days, 34.9% at 90 days, 32.4% at 200 days, and 30.4% at 365 days.

In vitro antibacterial activity

There was no reduction in the number of bacterial colonies in the pure titanium implants and the 0% iodine layer groups over 48 hours of incubation, and there was no significant difference between these two groups. However, all four iodine-supported implant groups (20%, 50%, 80%, and 100% iodine contents) showed significantly fewer *S. aureus* and *E. coli* colonies than the control groups, starting 2 hours after incubation ($p < 0.05$). Longer incubation periods resulted in greater decreases in the number of bacterial colonies in all iodine-supported implants, with complete disappearance of *S. aureus* and *E. coli* colonies by 6 hours in $\geq 80\%$ iodine groups (Figure 5). Implants with $\geq 20\%$ iodine completely inhibited *S. aureus* colonies after 24 hours and *E. coli* colonies after 6 hours (Figure 6 and 7).




Figure 5: *In vitro* antibacterial activity of titanium implants with different iodine contents on *S. aureus* after 6 hours of incubation. A: pure titanium implant, B: oxide layer with 0% iodine, C: 20%, D: 50%, E: 80%, and F: 100% iodine contents (where 100% corresponds to 13 $\mu\text{g}/\text{cm}^2$ iodine). The number of bacterial colonies was high in pure titanium implant and implants with an oxide layer with 0% iodine. In contrast, the number of colonies decreased gradually as the iodine content in implants increased.




Figure 6: *In vitro* antibacterial activity of titanium implants with different iodine contents on *S. aureus*. While pure titanium implants and 0% iodine implants had no antibacterial activity, implants with $\geq 20\%$ iodine showed a significant decrease in the number of colonies compared with the control group beginning at 2 hours ($p < 0.05$). Implants with 100% iodine exhibited significantly greater and faster decreases in the number of colonies compared to 20% iodine implants ($p < 0.05$). Implants with $\geq 20\%$ iodine showed complete inhibition of bacterial colonies after 24 hours.

The vertical axis is for bacterial colonies.




Figure 7: *In vitro* antibacterial activity of titanium implants with different iodine contents on *E. coli*. While pure titanium implants and 0% iodine implants had no antibacterial activity against *E. coli*, implants with $\geq 20\%$ iodine showed a significant decrease in the number of colonies compared with the control group beginning at 2 hours ($P < 0.05$). Implants with 80% and 100% iodine had significantly greater and faster decreases in the number of colonies compared to 20% and 50% iodine-supplemented implants ($p < 0.05$). Implants with $\geq 20\%$ iodine showed complete inhibition of bacterial colonies after 6 hours. The vertical axis is for bacterial colonies.

Discussion

The *in vitro* and *in vivo* experiments showed similar temporal patterns of attenuation of iodine contents in the implants. In the *in vitro* analysis, an initially rapid attenuation was followed by a more gradual decrease, with approximately 30% of the initial iodine maintained at 1 year. In the *in vivo* analysis, the iodine content in implants similarly decreased rapidly to 50–60% at 28 days, followed by a more gradual decline, with approximately 30% of initial iodine remaining in the implants at 1 year. We speculated that initial period of rapid decline can be attributed to the superficial pores, while the prolonged release is explained by the deep pores in the oxide layer.

The *in vitro* antibacterial activity of iodine-supported titanium implants was related to the iodine quantity in a dose-dependent and duration-dependent manner. While implants with high iodine contents ($\geq 80\%$) completely inhibited *S. aureus* and *E. coli* colonies by 6 hours, implants with lower iodine contents ($\geq 20\%$) also completely inhibited *S. aureus* colonies by 24 hours and *E. coli* colonies by 6 hours. These results suggest that implants with iodine contents of $\geq 20\%$ of the levels typically used in clinical applications possess sufficient antibacterial activity to prevent implant-related infections.

Based on the observed antibacterial activity and the temporal attenuation profile of the implants, currently used iodine-supported titanium implants possess adequate iodine to maintain antibacterial activity, even after 1 year of implantation. Further, the attenuation was similar at subcutaneous, intra-articular, and endo-osseous sites, suggesting that the implants effectively prevent infections in each of these locations.

The onset of implant-related infections can be early (1 to 3 months after implantation), delayed (within several months to 1 year after implantation), or late (beyond 1 year; infection carried by the blood from a non-surgical site infection) [24]. Therefore, it is crucial that the antibacterial implant retains antibacterial activity over long periods. Iodine-supported titanium implants are therefore particularly useful as they not only prevent early and late postoperative infections, but also likely prevent delayed infections. The long-term antibacterial activity of iodine-supported titanium implants is beneficial for implants used to treat fractures. Iodine-supported titanium implants are particularly

useful in cases of open fractures, which are associated with a higher risk of infection, and in cases with a risk of poor bone union, which may require plates and intramedullary nails for periods exceeding a year.

Several other surface treatment strategies, such as coating with antibiotics or silver ions, have also been used to impart antimicrobial properties to implant surfaces. However, many antibiotics act against a narrow spectrum of bacteria, and the emergence of resistant bacteria is an important risk [25]. Furthermore, the rapid release of antibiotics (80% in 60 minutes) limits their ability to effectively prevent infections over long periods [26]. Silver-coated orthopedic implants also show a rapid release of silver ions in the first three days, with only about 30% remaining after 14 days and antibacterial activity lasting for about 28 days [27,28]. Furthermore, toxicity resulting from high silver ion concentrations *in vivo* remains a concern. In particular, toxicity to bone cells and the accumulation of silver ions in the kidney, spleen, liver, and brain have been reported [29-32].

Iodine-supported titanium implants offer distinct advantages over implants with antibiotics and silver ion surface modifications. First, iodine has a wide antibacterial spectrum, is active against bacteria, viruses, fungi, and spores, and does not induce resistance in bacteria. Second, iodine has proven safety as a disinfectant and contrast agent [33]. Third, iodine-supported titanium implants possess long-term antibacterial activity that lasts longer than 1 year. In our previous *in vitro* and clinical studies, we demonstrated that iodine-supported titanium implants have sufficient antibacterial activity, but no cytotoxicity or deleterious effects on thyroid function [18,33]. We have also shown that iodine-supported titanium implants possess similar high bone affinity and osteoconductive properties to those of pure titanium implants [18].

This study had a few limitations. First, to examine antibacterial activity, we prepared titanium implants with low iodine contents to approximate the attenuated iodine content in implants after various periods. However, the *in vivo* release profile and antibacterial activity of these low-iodine-content titanium implants may differ from those of actual implants. Second, a relatively small number of implants were used per group. However, low variability within groups and repeated measurements from the same specimens enabled us to detect statistically significant differences.

In summary, iodine-supported titanium implants retain approximately 30% of the iodine content at 1 year in both *in vitro* and *in vivo* conditions. As implants with iodine contents of $\geq 20\%$ demonstrated sufficient antibacterial activity, current iodine-supported titanium implants can be expected to possess adequate antibacterial activity to prevent implant-related infections, even after one year of implantation, and should be strongly considered for orthopedic applications.

Acknowledgement

This work was supported by JSPS KAKENHI Grant Number 25670640. We would like to express our sincere gratitude to late Dr. Tohru Shimizu, Professor at Department of Bacteriology, Kanazawa University, for sharing his pearls of wisdom with us during the course of this research. Editorial support, in the form of medical writing based on authors' detailed directions, collating author comments, copyediting, fact checking, and referencing, was provided by Cactus Communications.

References

1. Bauer TW, Schils J (1999) The pathology of total joint arthroplasty. I. Mechanisms of implant fixation. *Skeletal Radiol* 28: 423-432.
2. Bauer TW, Schils J (1999) The pathology of total joint arthroplasty II Mechanisms of implant failure. *Skeletal Radiol* 28: 483-497.
3. Periti P, Mini E, Mosconi G (1998) Antimicrobial prophylaxis in orthopaedic surgery: the role of teicoplanin. *J Antimicrob Chemother* 41: 329-340.

4. Periti P, Stringa G, Mini E (1999) Comparative multicenter trial of teicoplanin versus cefazolin for antimicrobial prophylaxis in prosthetic joint implant surgery. Italian Study Group for Antimicrobial Prophylaxis in Orthopedic Surgery. *Eur J Clin Microbiol Infect Dis* 18: 113-119.
5. Sugarman B, Young EJ (1989) Infections associated with prosthetic devices: magnitude of the problem. *Infect Dis Clin North Am* 3: 187-198.
6. Kurtz SM, Lau E, Schmier J, Ong KL, Zhao K, et al. (2008) Infection burden for hip and knee arthroplasty in the United States. *J Arthroplasty* 23: 984-991.
7. Kurtz SM, Ong KL, Lau E, Bozic KJ, Berry D, et al. (2010) Prosthetic joint infection risk after TKA in the Medicare population. *Clin Orthop Relat Res* 468: 52-56.
8. Peersman G, Laskin R, Davis J, Peterson M (2001) Infection in total knee replacement: a retrospective review of 6489 total knee replacements. *Clin Orthop Relat Res* : 15-23.
9. Morii T, Yabe H, Morioka H, Beppu Y, Chuman H, et al. (2010) Postoperative deep infection in tumor endoprosthesis reconstruction around the knee. *J Orthop Sci* 15: 331-339.
10. Morii T, Morioka H, Ueda T, Araki N, Hashimoto N, et al. (2013) Deep infection in tumor endoprosthesis around the knee: a multi-institutional study by the Japanese musculoskeletal oncology group. *BMC Musculoskelet Disord* 14: 51.
11. Cen L, Neoh KG, Kang ET (2004) Antibacterial activity of cloth functionalized with N-alkylated poly(4-vinylpyridine). *J Biomed Mater Res A* 71: 70-80.
12. Yoshinari M, Oda Y, Kato T, Okuda K (2001) Influence of surface modifications to titanium on antibacterial activity in vitro. *Biomaterials* 22: 2043-2048.
13. Shirai T, Tsuchiya H, Shimizu T, Ohtani K, Zen Y, et al. (2009) Prevention of pin tract infection with titanium-copper alloys. *J Biomed Mater Res B Appl Biomater* 91: 373-380.
14. Harris LG, Mead L, Muller-Oberlander E, Richards RG (2006) Bacteria and cell cytocompatibility studies on coated medical grade titanium surfaces. *J Biomed Mater Res A* 78: 50-58.
15. Nohr RS, Macdonald JG (1994) New biomaterials through surface segregation phenomenon: new quaternary ammonium compounds as antibacterial agents. *J Biomater Sci Polym Ed* 5: 607-619.
16. Kinnari TJ, Peltonen LI, Kuusela P, Kivilahti J, Könönen M, et al. (2005) Bacterial adherence to titanium surface coated with human serum albumin. *Otol Neurotol* 26: 380-384.
17. Ewald A, Glückermann SK, Thull R, Gbureck U (2006) Antimicrobial titanium/silver PVD coatings on titanium. *Biomed Eng Online* 5: 22.
18. Shirai T, Shimizu T, Ohtani K, Zen Y, Takaya M, et al. (2011) Antibacterial iodine-supported titanium implants. *Acta Biomater* 7: 1928-1933.
19. Hashimoto K, Takaya M, Maejima A, Saruwatari K, Hirata M, et al. (1999) Antimicrobial characteristics of anodic oxidation coating of aluminum impregnated with iodine compound. *Inorg Mater* 6: 457-462.
20. Shirai T, Watanabe K, Matsubara H, Nomura I, Fujiwara H, et al. (2014) Prevention of pin tract infection with iodine-supported titanium pins. *J Orthop Sci* 19: 598-602.
21. Shirai T, Tsuchiya H, Nishida H, Yamamoto N, Watanabe K, et al. (2014) Antimicrobial megaprotheses supported with iodine. *J Biomater Appl* 29: 617-623.
22. Hayashi H, Murakami H, Demura S, Kato S, Yoshioka K, et al. (2015) Surgical site infection after total en bloc spondylectomy: risk factors and the preventive new technology. *Spine J* 15: 132-137.
23. Demura S, Murakami H, Shirai T, Kato S, Yoshioka K, et al. (2015) Surgical treatment for pyogenic vertebral osteomyelitis using iodine-supported spinal instruments: initial case series of 14 patients. *Eur J Clin Microbiol Infect Dis* 34: 261-266.
24. Vu DL, Uçkay I, Gonzalez A, Rohner P, Hoffmeyer P, et al. (2016) Factors related to outcome of early and delayed prosthetic joint infections. *J Infect* 72: 255-257.
25. Oduwale KO, Glynn AA, Monoly DC, Murray D, Rowe S, et al. (2010) Antibiofilm activity of sub-inhibitory povidone-iodine concentrations against *Staphylococcus epidermidis* and *Staphylococcus aureus*. *J Orthop Res* 28: 1252-1256.
26. Stigter M, de Groot K, Layrolle P (2002) Incorporation of tobramycin into biomimetic hydroxyapatite coating on titanium. *Biomaterials* 23: 4143-4153.
27. Zhang L, Zhang L, Yang Y, Zhang W, Lv H, et al. (2016) Inhibitory effect of super-hydrophobicity on silver release and antibacterial properties of super-hydrophobic Ag/TiO₂ nanotubes. *J Biomed Mater Res B Appl Biomater* 104: 1004-1012.
28. Cheng H, Li Y, Huo K, Gao B, Xiong W (2014) Long-lasting in vivo and in vitro antibacterial ability of nanostructured titania coating incorporated with silver nanoparticles. *J Biomed Mater Res A* 102: 3488-3499.
29. Massè A, Bruno A, Bosetti M, Biasibetti A, Cannas M, et al. (2000) Prevention of pin track infection in external fixation with silver coated pins: clinical and microbiological results. *J Biomed Mater Res* 53: 600-604.
30. Hards J, Ahrens H, Gebert C, Streitburger A, Buerger H, et al. (2007) Lack of toxicological side-effects in silver-coated megaprotheses in humans. *Biomaterials* 28: 2869-2875.
31. Albers CE, Hofstetter W, Siebenrock KA, Landmann R, Klenke FM (2013) In vitro cytotoxicity of silver nanoparticles on osteoblasts and osteoclasts at antibacterial concentrations. *Nanotoxicology* 7: 30-36.
32. Gosheger G, Hards J, Ahrens H, Streitburger A, Buerger H, et al. (2004) Silver-coated megaendoprotheses in a rabbit model—an analysis of the infection rate and toxicological side effects. *Biomaterials* 25: 5547-5556.
33. Tsuchiya H, Shirai T, Nishida H, Murakami H, Kabata T, et al. (2012) Innovative antimicrobial coating of titanium implants with iodine. *J Orthop Sci* 17: 595-604.

Citation: Kato T, Shirai T, Yamamoto N, Nishida H, Hayashi K, et al. (2016) Temporal Attenuation of Iodine Content and its Effect on the Antibacterial Activity of Iodine-Supported Titanium Implants. *J Microb Biochem Technol* 8: 285-289. doi: [10.4172/1948-5948.1000298](https://doi.org/10.4172/1948-5948.1000298)

OMICS International: Open Access Publication Benefits & Features

Unique features:

- Increased global visibility of articles through worldwide distribution and indexing
- Showcasing recent research output in a timely and updated manner
- Special issues on the current trends of scientific research

Special features:

- 700+ Open Access Journals
- 50,000+ editorial team
- Rapid peer review process
- Quality and quick editorial, review and publication processing
- Indexing at major indexing services
- Sharing Option: Social Networking Enabled for better prominence and citations
- Authors, Reviewers and Editors rewarded with online Scientific Credits
- Best discounts for your subsequent articles

Submit your manuscript at: <http://www.editorialmanager.com/jmbt>